

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2824

Name: RITTER 1  
Sterilizer ID: V2178198  
01/30/2023 11:01 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 186.6 F 0.0  
2:00 188.7 F 3.9  
4:00 194.5 F 8.2  
6:00 211.1 F 13.1  
8:00 226.5 F 7.2  
10:00 238.6 F 12.3  
12:00 235.6 F 9.0  
14:00 238.3 F 9.9  
16:00 246.5 F 13.1  
18:00 258.3 F 19.6  
20:00 269.0 F 26.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.0  
1:00 271.9 F 29.1  
2:00 272.6 F 29.2  
3:00 271.9 F 28.8  
4:00 272.7 F 29.5

Min 270.9 F 27.9  
Max 273.5 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.1 F 0.0  
5:00 213.7 F 0.0  
10:00 227.5 F 0.0  
15:00 232.1 F 0.0  
20:00 227.6 F 0.0  
25:00 226.1 F 0.0  
30:00 227.2 F 0.0  
Drying Complete

FILLING: 1:11  
HEATING: 21:09  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 00:59:37  
CYCLE COMPLETE

01/30/2023 12:00 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_