
Midmark M11 - v1.0.0
Total Cycles: 2770

Name: RITTER 1
Sterilizer ID: V2178198
01/11/2023 05:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.7 F 0.0
2:00 195.4 F 1.2
4:00 206.9 F 6.6
6:00 222.6 F 12.3
8:00 231.5 F 8.3
10:00 240.5 F 11.6
12:00 243.7 F 12.3
14:00 242.4 F 11.2
16:00 256.0 F 18.3
18:00 267.7 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.1
1:00 272.3 F 28.9
2:00 271.9 F 28.3
3:00 271.8 F 28.5
4:00 272.3 F 28.7

Min 271.4 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.5 F 0.0
5:00 224.0 F 0.0
10:00 232.5 F 0.0
15:00 232.0 F 0.0
20:00 226.4 F 0.0
25:00 225.4 F 0.0
30:00 224.1 F 0.0

Drying Complete

FILLING: 1:45
HEATING: 19:12
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:10
CYCLE COMPLETE

01/11/2023 06:05 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____