
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
1.4uS / 0.8ppm
CYCLE NUMBER 000809
14:06 17/11/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:22
PRESSURIZING 2:28
STERILIZING 3:24
136.0 C 228kPa 3:24
136.2 C 231kPa 3:54
136.3 C 231kPa 4:24
136.5 C 233kPa 4:54
136.3 C 232kPa 5:24
136.4 C 232kPa 5:54
136.4 C 233kPa 6:24
136.1 C 231kPa 6:54
136.1 C 231kPa 6:54
VENTING 6:54
AIR DRYING 7:41
CYCLE COMPLETE 21:26

Digital Signature #
3A56E0535C85D22C
