
Midmark M11 - v1.0.0
Total Cycles: 3608

Name: RITTER 1
Sterilizer ID: V2178198
08/02/2023 12:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.2 F 0.0
2:00 184.5 F 1.5
4:00 192.1 F 5.1
6:00 214.5 F 9.6
8:00 221.6 F 8.7
10:00 229.7 F 8.0
12:00 240.0 F 12.6
14:00 236.6 F 9.5
16:00 240.7 F 9.9
18:00 246.6 F 13.4
20:00 257.5 F 19.2
22:00 267.3 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.7 F 28.9
2:00 271.8 F 28.2
3:00 272.0 F 28.7
4:00 272.4 F 28.8

Min 271.3 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.5 F 0.0
5:00 216.4 F 0.0
10:00 219.9 F 0.0
15:00 222.0 F 0.0
20:00 218.9 F 0.0
25:00 216.9 F 0.0
30:00 216.4 F 0.0
Drying Complete

FILLING: 1:27
HEATING: 23:34
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:02:34
CYCLE COMPLETE

08/02/2023 01:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____