

-----  
Midmark M11 - v1.0.5  
Total Cycles: 507

Name: RITTER2  
Sterilizer ID: V2313228  
11/01/2021 06:42 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 72.2 F 0.0  
2:00 73.3 F 0.0  
4:00 74.8 F 0.0  
6:00 86.1 F 0.1  
8:00 124.1 F 0.2  
10:00 169.7 F 0.3  
12:00 200.7 F 3.2  
14:00 214.5 F 6.4  
16:00 225.3 F 10.0  
18:00 230.7 F 10.3  
20:00 236.4 F 11.6  
22:00 235.7 F 10.0  
24:00 235.7 F 8.9  
26:00 249.8 F 16.0  
28:00 261.8 F 22.8  
30:00 272.6 F 30.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.2  
1:00 272.5 F 29.6  
2:00 272.3 F 29.9  
3:00 271.9 F 29.2  
4:00 272.8 F 30.0

Min 271.5 F 29.1  
Max 272.8 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.1 F 0.0  
5:00 194.9 F 0.0  
10:00 188.3 F 0.0  
15:00 186.6 F 0.0  
20:00 184.1 F 0.0  
25:00 190.9 F 0.0  
30:00 156.0 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 30:23  
STERILIZING: 4:00  
VENTING: 3:03  
DRYING: 30:00  
TOTAL CYCLE: 01:08:19  
CYCLE COMPLETE

11/01/2021 08:01 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01