

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1786

Name: RITTER2  
Sterilizer ID: V2313228  
11/29/2022 03:03 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 162.5 F 0.0  
2:00 156.6 F 0.1  
4:00 184.8 F 0.8  
6:00 210.9 F 4.8  
8:00 226.2 F 9.8  
10:00 232.8 F 8.4  
12:00 240.3 F 12.4  
14:00 240.4 F 11.0  
16:00 239.8 F 10.3  
18:00 254.1 F 17.6  
20:00 266.1 F 24.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.6  
1:00 272.5 F 29.1  
2:00 271.5 F 28.8  
3:00 272.6 F 29.2  
4:00 271.5 F 28.5

Min 271.4 F 28.5  
Max 272.8 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.0 F 0.0  
5:00 224.4 F 0.0  
10:00 233.1 F 0.0  
15:00 228.0 F 0.0  
20:00 226.6 F 0.0  
25:00 227.2 F 0.0  
30:00 225.5 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 21:35  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:59:49  
CYCLE COMPLETE

11/29/2022 04:02 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_