
Midmark M11 - v1.0.0
Total Cycles: 4109

Name: RITTER 1
Sterilizer ID: V2178198
12/12/2023 12:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 94.2 F 0.0
2:00 93.9 F 0.5
4:00 98.4 F 1.8
6:00 123.0 F 4.1
8:00 146.9 F 6.7
10:00 172.4 F 10.1
12:00 199.0 F 10.6
14:00 209.1 F 5.6
16:00 222.5 F 9.2
18:00 233.5 F 13.2
20:00 233.5 F 8.9
22:00 243.8 F 12.3
24:00 242.2 F 11.6
26:00 253.2 F 17.3
28:00 263.6 F 23.3
30:00 272.4 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.1
1:00 272.9 F 29.3
2:00 272.1 F 29.5
3:00 272.2 F 29.2
4:00 272.5 F 29.7

Min 271.5 F 28.9
Max 273.0 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.6 F 0.0
5:00 227.5 F 0.0
10:00 217.4 F 0.0
15:00 212.1 F 0.0
20:00 206.1 F 0.0
25:00 215.5 F 0.0
30:00 223.4 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 31:01
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:09:44
CYCLE COMPLETE

12/12/2023 01:23 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____