
Midmark M11 - v1.0.0
Total Cycles: 1186

Name: RITTER 1
Sterilizer ID: V2178198
09/20/2021 09:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.8 F 0.0
2:00 180.0 F 2.6
4:00 189.8 F 6.7
6:00 209.5 F 11.6
8:00 219.9 F 5.2
10:00 232.7 F 10.1
12:00 235.3 F 8.6
14:00 243.7 F 12.6
16:00 242.1 F 11.2
18:00 254.5 F 17.5
20:00 265.3 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 272.7 F 28.8
2:00 271.8 F 28.1
3:00 272.4 F 28.8
4:00 272.4 F 28.5

Min 271.6 F 28.0
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.7 F 0.0
5:00 212.1 F 0.0
10:00 179.5 F 0.0
15:00 168.4 F 0.0
20:00 174.0 F 0.0
25:00 177.0 F 0.0
30:00 206.2 F 0.0
Drying Complete

FILLING: 1:31
HEATING: 21:52
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:00:41
CYCLE COMPLETE

09/20/2021 10:21 AM
ACCEPTED: X
REJECTED:
Operator: U01