
Midmark M11 - v1.0.0
Total Cycles: 2956

Name: RITTER 1
Sterilizer ID: V2178198
02/28/2023 03:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 151.0 F 0.0
2:00 151.5 F 0.6
4:00 175.9 F 3.7
6:00 201.8 F 8.5
8:00 218.6 F 9.8
10:00 226.8 F 8.4
12:00 238.3 F 12.3
14:00 239.9 F 11.4
16:00 238.4 F 9.9
18:00 252.6 F 16.9
20:00 264.8 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.6 F 29.1
2:00 271.5 F 28.7
3:00 272.5 F 29.2
4:00 271.6 F 28.5

Min 271.5 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.0 F 0.0
5:00 213.6 F 0.0
10:00 229.9 F 0.0
15:00 232.3 F 0.0
20:00 227.6 F 0.0
25:00 231.2 F 0.0
30:00 229.2 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 21:47
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:00:01
CYCLE COMPLETE

02/28/2023 04:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____