

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 007796  
12:51 23/08/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:46  
PRESSURIZING 2:48  
STERILIZING 3:51  
136.0 C 220kPa 3:51  
136.4 C 222kPa 4:21  
136.0 C 218kPa 4:51  
136.4 C 222kPa 5:21  
136.0 C 222kPa 5:51  
136.4 C 224kPa 6:21  
136.4 C 224kPa 6:51  
136.4 C 223kPa 7:21  
136.4 C 223kPa 7:21  
VENTING 7:21  
AIR DRYING 8:11  
CYCLE COMPLETE 38:11

-----  
Digital Signature #  
ECD6451292D0AB64  
-----