

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4848

Name: RITTER 1  
Sterilizer ID: V2178198  
08/23/2024 11:01 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 115.5 F 0.0  
2:00 119.3 F 0.2  
4:00 145.9 F 0.9  
6:00 174.6 F 2.5  
8:00 193.7 F 5.8  
10:00 211.8 F 9.4  
12:00 222.0 F 11.7  
14:00 226.4 F 7.5  
16:00 236.4 F 12.1  
18:00 235.2 F 8.8  
20:00 241.8 F 10.9  
22:00 245.2 F 12.8  
24:00 256.8 F 18.8  
26:00 266.9 F 25.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.1  
1:00 272.2 F 28.9  
2:00 272.1 F 28.6  
3:00 271.3 F 28.0  
4:00 271.5 F 28.5

Min 271.1 F 28.0  
Max 272.4 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.3 F 0.0  
5:00 209.5 F 0.0  
10:00 207.1 F 0.0  
15:00 217.3 F 0.0  
20:00 238.3 F 0.0  
25:00 229.2 F 0.0  
30:00 226.9 F 0.0  
Drying Complete

FILLING: 1:16  
HEATING: 27:40  
STERILIZING: 4:00  
VENTING: 3:43  
DRYING: 30:00  
TOTAL CYCLE: 01:06:39  
CYCLE COMPLETE

08/23/2024 12:07 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_