
Midmark M11 - v1.0.5
Total Cycles: 4119

Name: RITTER2
Sterilizer ID: V2313228
11/14/2024 01:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 137.0 F 0.0
2:00 136.7 F 0.2
4:00 156.8 F 2.8
6:00 182.3 F 6.2
8:00 199.2 F 9.8
10:00 213.9 F 9.5
12:00 223.9 F 7.1
14:00 234.8 F 11.4
16:00 232.8 F 8.4
18:00 243.0 F 12.2
20:00 244.6 F 12.9
22:00 256.4 F 19.1
24:00 266.8 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.7 F 29.5
2:00 271.5 F 28.7
3:00 272.4 F 29.6
4:00 272.4 F 29.1

Min 271.2 F 28.6
Max 272.9 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.2 F 0.0
5:00 233.9 F 0.0
10:00 220.7 F 0.0
15:00 215.8 F 0.0
20:00 217.2 F 0.0
25:00 226.1 F 0.0
30:00 227.5 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 25:35
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:04:19
CYCLE COMPLETE

11/14/2024 02:40 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____