

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4168

Name: RITTER 1  
Sterilizer ID: V2178198  
12/29/2023 10:33 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.2 F 0.0  
2:00 71.9 F 0.3  
4:00 83.8 F 1.5  
6:00 120.3 F 4.1  
8:00 150.2 F 7.4  
10:00 172.7 F 11.3  
12:00 205.7 F 2.6  
14:00 219.4 F 6.5  
16:00 229.3 F 10.4  
18:00 235.1 F 8.2  
20:00 239.3 F 10.9  
22:00 235.5 F 8.3  
24:00 249.7 F 14.9  
26:00 261.6 F 21.6  
28:00 272.3 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.2  
1:00 271.3 F 28.1  
2:00 271.5 F 28.1  
3:00 271.7 F 28.1  
4:00 271.9 F 28.2

Min 271.3 F 28.0  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 234.1 F 0.0  
Dry Cycle Aborted

FILLING: 1:03  
HEATING: 28:21  
STERILIZING: 4:00  
VENTING: 2:58  
DRYING: 1:38  
TOTAL CYCLE: 00:38:00

12/29/2023 11:10 AM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_