

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002053  
11:23 18/09/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:01  
PRESSURIZING 2:06  
STERILIZING 3:06  
136.0 C 220kPa 3:06  
135.7 C 220kPa 3:36  
136.2 C 223kPa 4:06  
136.1 C 221kPa 4:36  
136.4 C 224kPa 5:06  
136.4 C 224kPa 5:36  
136.4 C 224kPa 6:06  
136.3 C 224kPa 6:36  
136.3 C 224kPa 6:36  
VENTING 6:36  
AIR DRYING 7:27  
CYCLE COMPLETE 37:27

-----  
Digital Signature #  
9C3EB398E80538EC  
-----