
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002929
11:37 11/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:38
PRESSURIZING 3:38
STERILIZING 4:50
136.4 C 223kPa 4:50
136.3 C 223kPa 5:20
136.2 C 222kPa 5:50
136.4 C 219kPa 6:20
136.4 C 224kPa 6:50
136.2 C 221kPa 7:20
136.4 C 224kPa 7:50
136.1 C 219kPa 8:20
136.1 C 219kPa 8:20
VENTING 8:20
AIR DRYING 9:15
CYCLE COMPLETE 9:17

Digital Signature #
BCAC81B7CC3E7787
