

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5026

Name: RITTER 1  
Sterilizer ID: V2178198  
10/25/2024 01:30 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 156.2 F 0.0  
2:00 156.0 F 0.3  
4:00 171.4 F 1.4  
6:00 194.5 F 5.3  
8:00 209.0 F 9.3  
10:00 230.0 F 10.8  
12:00 230.6 F 8.5  
14:00 241.3 F 13.3  
16:00 239.1 F 10.6  
18:00 237.7 F 8.6  
20:00 250.3 F 15.3  
22:00 261.8 F 21.8  
24:00 272.4 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.2  
1:00 272.5 F 29.0  
2:00 272.2 F 28.5  
3:00 271.4 F 28.1  
4:00 271.7 F 28.6

Min 271.2 F 28.0  
Max 272.5 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.3 F 0.0  
5:00 214.2 F 0.0  
10:00 208.3 F 0.0  
15:00 212.6 F 0.0  
20:00 228.4 F 0.0  
25:00 232.7 F 0.0  
30:00 230.2 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 24:32  
STERILIZING: 4:00  
VENTING: 3:47  
DRYING: 30:00  
TOTAL CYCLE: 01:03:10  
CYCLE COMPLETE

10/25/2024 02:33 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_