

-----  
Mississauga, ON  
Canada  
-----

STATIM 5000 S501R751  
SN 511020J00004  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000832  
11:30 31/03/2022  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:26  
PRESSURIZING 6:13  
STERILIZING 7:38  
136.0 C 220kPa 7:38  
136.4 C 223kPa 8:08  
136.4 C 223kPa 8:38  
136.4 C 223kPa 9:08  
136.4 C 223kPa 9:38  
136.2 C 222kPa 10:08  
136.4 C 223kPa 10:38  
136.2 C 222kPa 11:08  
136.2 C 222kPa 11:08  
VENTING 11:08  
AIR DRYING 12:13  
CYCLE COMPLETE 33:39

-----  
Digital Signature #  
D70F52CC2AE011A6  
-----