

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000200  
10:14 21/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:37  
PRESSURIZING 3:43  
STERILIZING 5:27  
136.4 C 232kPa 5:27  
136.3 C 231kPa 5:57  
136.3 C 232kPa 6:27  
136.3 C 232kPa 6:57  
136.3 C 231kPa 7:27  
136.3 C 233kPa 7:57  
136.3 C 231kPa 8:27  
136.3 C 232kPa 8:57  
136.3 C 232kPa 8:57  
VENTING 8:57  
AIR DRYING 10:00  
CYCLE COMPLETE 18:50  
-----

Digital Signature #  
17C20B3D6599D0E6  
-----