
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001208
9:36 11/08/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:19
PRESSURIZING 4:20
STERILIZING 5:39
136.0 C 220kPa 5:39
136.3 C 224kPa 6:09
136.2 C 223kPa 6:39
136.2 C 223kPa 7:09
136.3 C 224kPa 7:39
136.1 C 222kPa 8:09
136.3 C 223kPa 8:39
136.4 C 224kPa 9:09
136.4 C 224kPa 9:09
VENTING 9:09
AIR DRYING 10:06
CYCLE COMPLETE 25:07

Digital Signature #
F1FA327CB1BF4859
