

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4390

Name: RITTER2  
Sterilizer ID: V2313228  
02/28/2025 11:52 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 154.5 F 0.0  
2:00 148.9 F 1.6  
4:00 150.3 F 4.4  
6:00 177.0 F 8.1  
8:00 191.7 F 12.5  
10:00 205.9 F 5.3  
12:00 219.3 F 9.6  
14:00 229.9 F 10.5  
16:00 234.7 F 11.3  
18:00 235.4 F 9.6  
20:00 247.7 F 15.5  
22:00 258.4 F 21.4  
24:00 268.0 F 27.9

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 30.1  
1:00 271.8 F 29.9  
2:00 271.9 F 30.3  
3:00 272.3 F 30.7  
4:00 272.8 F 31.1

Min 271.7 F 29.9  
Max 273.4 F 31.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.3 F 0.0  
5:00 204.5 F 0.0  
10:00 208.7 F 0.0  
15:00 207.0 F 0.0  
20:00 205.7 F 0.0  
25:00 208.6 F 0.0  
30:00 210.6 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 25:20  
STERILIZING: 4:00  
VENTING: 3:59  
DRYING: 30:00  
TOTAL CYCLE: 01:04:15  
CYCLE COMPLETE

02/28/2025 12:56 PM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_