

-----  
Midmark M11 - v1.0.0  
Total Cycles: 993

Name: RITTER 1  
Sterilizer ID: V2178198  
07/12/2021 08:58 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 188.6 F 0.0  
2:00 190.6 F 2.3  
4:00 195.9 F 6.6  
6:00 216.6 F 11.8  
8:00 224.3 F 7.3  
10:00 237.8 F 12.8  
12:00 237.6 F 10.2  
14:00 237.7 F 8.0  
16:00 249.5 F 15.0  
18:00 260.9 F 21.4  
20:00 271.6 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.6 F 28.9  
2:00 271.6 F 28.6  
3:00 272.3 F 28.9  
4:00 271.4 F 28.3

Min 271.4 F 28.2  
Max 272.9 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.9 F 0.0  
5:00 219.3 F 0.0  
10:00 233.2 F 0.0  
15:00 221.8 F 0.0  
20:00 217.2 F 0.0  
25:00 214.2 F 0.0  
30:00 211.5 F 0.0  
Drying Complete

FILLING: 1:35  
HEATING: 20:36  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:59:18  
CYCLE COMPLETE

07/12/2021 09:58 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01