
Midmark M11 - v1.0.0
Total Cycles: 3234

Name: RITTER 1
Sterilizer ID: V2178198
05/15/2023 02:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.5 F 0.0
2:00 185.6 F 1.5
4:00 200.2 F 6.4
6:00 218.4 F 12.1
8:00 227.4 F 7.4
10:00 239.7 F 12.7
12:00 238.7 F 10.3
14:00 237.3 F 8.1
16:00 250.0 F 15.2
18:00 261.7 F 21.8
20:00 272.1 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.5 F 29.1
2:00 272.1 F 28.6
3:00 271.3 F 28.2
4:00 272.2 F 29.0

Min 271.2 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.4 F 0.0
5:00 216.6 F 0.0
10:00 236.9 F 0.0
15:00 244.3 F 0.0
20:00 235.7 F 0.0
25:00 224.7 F 0.0
30:00 222.6 F 0.0
Drying Complete

FILLING: 1:34
HEATING: 20:35
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 00:59:28
CYCLE COMPLETE

05/15/2023 03:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____