
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001774
12:47 15/05/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:43
PRESSURIZING 2:49
STERILIZING 3:52
136.4 C 221kPa 3:52
135.9 C 220kPa 4:22
135.9 C 221kPa 4:52
136.4 C 225kPa 5:22
136.3 C 223kPa 5:52
136.2 C 220kPa 6:22
136.4 C 224kPa 6:52
136.4 C 224kPa 7:22
136.4 C 224kPa 7:22
VENTING 7:22
AIR DRYING 8:13
CYCLE COMPLETE 38:13

Digital Signature #
C1B8D651FFB86172
