
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001413
9:06 09/09/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:11
PRESSURIZING 5:47
STERILIZING 7:04
136.0 C 220kPa 7:04
136.0 C 220kPa 7:34
136.4 C 220kPa 8:04
136.5 C 223kPa 8:34
136.4 C 223kPa 9:04
136.1 C 220kPa 9:34
136.0 C 217kPa 10:04
136.4 C 223kPa 10:34
VENTING 10:34
AIR DRYING 11:40
CYCLE COMPLETE 11:46

Digital Signature #
D116AB712F616D45
