
Midmark M11 - v1.0.0
Total Cycles: 4393

Name: RITTER 1
Sterilizer ID: V2178198
03/21/2024 03:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 143.4 F 0.0
2:00 144.4 F 2.1
4:00 153.0 F 5.0
6:00 182.0 F 8.8
8:00 197.3 F 13.5
10:00 216.5 F 6.2
12:00 226.5 F 10.7
14:00 230.7 F 7.1
16:00 241.4 F 12.4
18:00 239.8 F 10.5
20:00 252.3 F 16.7
22:00 263.5 F 23.1
24:00 271.8 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.1 F 28.8
2:00 272.0 F 29.2
3:00 271.5 F 28.5
4:00 272.4 F 29.1

Min 271.2 F 28.4
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.8 F 0.0
5:00 213.7 F 0.0
10:00 222.3 F 0.0
15:00 227.8 F 0.0
20:00 233.1 F 0.0
25:00 230.7 F 0.0
30:00 227.8 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 24:09
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:02:55
CYCLE COMPLETE

03/21/2024 04:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____