

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000971  
9:17 14/12/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:56  
PRESSURIZING 3:56  
STERILIZING 5:02  
136.0 C 220kPa 5:02  
136.2 C 222kPa 5:32  
136.4 C 224kPa 6:02  
136.1 C 218kPa 6:32  
136.3 C 221kPa 7:02  
136.2 C 223kPa 7:32  
136.3 C 222kPa 8:02  
136.2 C 218kPa 8:32  
136.2 C 218kPa 8:32  
VENTING 8:32  
AIR DRYING 9:30  
CYCLE COMPLETE 9:32

-----  
Digital Signature #  
A7D0D6B3CAA9D9FB  
-----