
Midmark M11 - v1.0.5
Total Cycles: 2072

Name: RITTER2
Sterilizer ID: V2313228
02/21/2023 01:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.3 F 0.0
2:00 193.4 F 3.4
4:00 199.6 F 8.7
6:00 223.1 F 8.8
8:00 232.4 F 11.2
10:00 235.3 F 9.6
12:00 235.9 F 9.0
14:00 251.2 F 16.5
16:00 263.9 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.2
1:00 272.3 F 29.8
2:00 271.3 F 29.1
3:00 272.3 F 30.1
4:00 272.1 F 29.6

Min 271.1 F 29.0
Max 272.5 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.5 F 0.0
5:00 218.2 F 0.0
10:00 231.1 F 0.0
15:00 235.3 F 0.0
20:00 226.3 F 0.0
25:00 224.9 F 0.0
30:00 227.6 F 0.0
Drying Complete

FILLING: 1:51
HEATING: 17:55
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:57:03
CYCLE COMPLETE

02/21/2023 01:58 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____