

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1889

Name: RITTER2  
Sterilizer ID: V2313228  
01/03/2023 09:48 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 147.3 F 0.0  
2:00 145.6 F 0.6  
4:00 159.1 F 3.1  
6:00 185.2 F 7.5  
8:00 210.9 F 10.8  
10:00 228.0 F 9.4  
12:00 233.5 F 8.6  
14:00 243.7 F 12.2  
16:00 245.6 F 13.2  
18:00 259.2 F 20.6  
20:00 271.1 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.8  
1:00 272.3 F 28.8  
2:00 273.0 F 29.4  
3:00 272.2 F 29.2  
4:00 272.3 F 28.9

Min 271.7 F 28.6  
Max 273.1 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 222.8 F 0.0  
5:00 234.5 F 0.0  
Dry Cycle Aborted

FILLING: 0:52  
HEATING: 20:38  
STERILIZING: 4:00  
VENTING: 2:42  
DRYING: 8:29  
TOTAL CYCLE: 00:36:42

01/03/2023 10:24 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_