

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000582  
12:58 12/08/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:24  
PRESSURIZING 2:32  
STERILIZING 3:40  
136.0 C 228kPa 3:40  
136.4 C 233kPa 4:10  
136.4 C 233kPa 4:40  
136.4 C 233kPa 5:10  
136.4 C 231kPa 5:40  
136.4 C 233kPa 6:10  
136.4 C 233kPa 6:40  
136.3 C 233kPa 7:10  
136.3 C 233kPa 7:10  
VENTING 7:11  
AIR DRYING 8:04  
CYCLE COMPLETE 23:04  
-----

Digital Signature #  
7D43A35FF7EF37BA  
-----