
Midmark M11 - v1.0.5
Total Cycles: 4342

Name: RITTER2
Sterilizer ID: V2313228
02/07/2025 03:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 93.6 F 0.0
2:00 94.7 F 0.5
4:00 104.0 F 2.2
6:00 126.8 F 4.8
8:00 164.5 F 7.9
10:00 177.1 F 11.5
12:00 208.8 F 3.8
14:00 217.0 F 7.1
16:00 223.7 F 10.7
18:00 230.7 F 7.7
20:00 238.4 F 11.4
22:00 236.3 F 9.2
24:00 248.1 F 14.9
26:00 258.7 F 20.7
28:00 268.4 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.2
1:00 273.1 F 30.4
2:00 271.9 F 30.0
3:00 272.7 F 30.0
4:00 272.6 F 30.5

Min 270.9 F 29.1
Max 273.3 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.9 F 0.0
5:00 221.6 F 0.0
10:00 211.1 F 0.0
15:00 202.5 F 0.0
20:00 197.9 F 0.0
25:00 202.0 F 0.0
30:00 203.2 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 29:15
STERILIZING: 4:00
VENTING: 3:52
DRYING: 30:00
TOTAL CYCLE: 01:08:04
CYCLE COMPLETE

02/07/2025 04:22 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____