
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000090
11:39 28/08/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:29
PRESSURIZING 3:32
STERILIZING 4:39
136.0 C 220kPa 4:39
136.2 C 225kPa 5:09
136.4 C 226kPa 5:39
136.3 C 224kPa 6:09
136.4 C 225kPa 6:39
136.4 C 226kPa 7:09
136.1 C 223kPa 7:39
136.3 C 223kPa 8:09
136.3 C 223kPa 8:09
VENTING 8:09
AIR DRYING 9:04
CYCLE COMPLETE 24:04

Digital Signature #
8844A809E34E0CB1
