
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002598
12:08 24/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:27
PRESSURIZING 2:31
STERILIZING 3:35
136.0 C 222kPa 3:35
136.4 C 225kPa 4:05
136.4 C 224kPa 4:35
136.3 C 225kPa 5:05
136.4 C 226kPa 5:35
136.4 C 225kPa 6:05
136.4 C 225kPa 6:35
136.2 C 224kPa 7:05
136.2 C 224kPa 7:05
VENTING 7:05
AIR DRYING 8:04
CYCLE COMPLETE 11:11

Digital Signature #
35EEEEEF0C441BD5
