
Midmark M11 - v1.0.0
Total Cycles: 3085

Name: RITTER 1
Sterilizer ID: V2178198
04/10/2023 04:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.9 F 0.0
2:00 175.7 F 3.4
4:00 188.7 F 7.9
6:00 218.5 F 13.0
8:00 224.1 F 8.2
10:00 236.1 F 13.1
12:00 238.3 F 10.7
14:00 235.8 F 8.6
16:00 250.1 F 15.5
18:00 262.2 F 22.2
20:00 272.7 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.9 F 29.3
2:00 271.8 F 28.5
3:00 271.8 F 29.0
4:00 272.8 F 29.1

Min 271.2 F 28.4
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.2 F 0.0
5:00 235.6 F 0.0
10:00 232.8 F 0.0
15:00 227.4 F 0.0
20:00 229.7 F 0.0
25:00 231.5 F 0.0
30:00 228.1 F 0.0
Drying Complete

FILLING: 1:39
HEATING: 20:25
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:13
CYCLE COMPLETE

04/10/2023 05:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____