
Midmark M11 - v1.0.0
Total Cycles: 1181

Name: RITTER 1
Sterilizer ID: V2178198
09/16/2021 11:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.1 F 0.0
2:00 187.1 F 3.3
4:00 200.9 F 8.1
6:00 220.6 F 11.7
8:00 229.2 F 8.4
10:00 240.5 F 12.1
12:00 241.6 F 11.3
14:00 238.9 F 9.8
16:00 253.1 F 16.7
18:00 265.0 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 272.7 F 28.7
2:00 271.5 F 28.0
3:00 272.8 F 28.9
4:00 271.9 F 28.2

Min 271.5 F 27.9
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.6 F 0.0
5:00 196.3 F 0.0
10:00 199.2 F 0.0
15:00 199.4 F 0.0
20:00 199.0 F 0.0
25:00 197.7 F 0.0
30:00 197.7 F 0.0
Drying Complete

FILLING: 1:44
HEATING: 19:49
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:58:39
CYCLE COMPLETE

09/16/2021 01:02 PM
ACCEPTED: X
REJECTED:
Operator: U02