
Midmark M11 - v1.0.5
Total Cycles: 3047

Name: RITTER2
Sterilizer ID: V2313228
10/24/2023 08:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.1 F 0.0
2:00 71.3 F 0.0
4:00 82.8 F 0.1
6:00 105.8 F 1.1
8:00 140.6 F 3.7
10:00 164.2 F 7.2
12:00 206.8 F 10.9
14:00 217.6 F 4.7
16:00 220.8 F 7.2
18:00 231.1 F 11.1
20:00 229.2 F 6.7
22:00 242.7 F 12.3
24:00 239.8 F 10.5
26:00 252.6 F 16.8
28:00 263.7 F 23.3
30:00 272.2 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 273.0 F 29.4
2:00 272.3 F 29.8
3:00 271.8 F 29.0
4:00 272.6 F 29.7

Min 271.4 F 29.0
Max 273.1 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.1 F 0.0
5:00 220.3 F 0.0
10:00 212.7 F 0.0
15:00 173.1 F 0.0
20:00 170.3 F 0.0
25:00 168.1 F 0.0
30:00 165.8 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:57
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:09:03
CYCLE COMPLETE

10/24/2023 09:20 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____