
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002004
12:02 10/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:39
PRESSURIZING 2:42
STERILIZING 3:48
136.4 C 225kPa 3:48
136.4 C 226kPa 4:18
136.4 C 222kPa 4:48
136.3 C 224kPa 5:18
136.4 C 226kPa 5:48
136.4 C 226kPa 6:18
136.3 C 224kPa 6:48
136.4 C 226kPa 7:18
136.4 C 226kPa 7:18
VENTING 7:18
AIR DRYING 8:16
CYCLE COMPLETE 10:15

Digital Signature #
995DF715882CD4C0
