
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000631
13:17 03/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:47
PRESSURIZING 3:53
STERILIZING 5:05
136.5 C 232kPa 5:05
136.2 C 231kPa 5:35
136.2 C 230kPa 6:05
136.3 C 232kPa 6:35
136.4 C 230kPa 7:05
136.4 C 232kPa 7:35
136.2 C 231kPa 8:05
136.2 C 231kPa 8:35
136.2 C 231kPa 8:35
VENTING 8:36
AIR DRYING 9:30
CYCLE COMPLETE 24:30

Digital Signature #
94F9105D49F8E96F
