
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 007526
11:33 10/06/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:47
PRESSURIZING 2:49
STERILIZING 3:50
136.0 C 220kPa 3:50
136.2 C 222kPa 4:20
136.2 C 223kPa 4:50
136.4 C 224kPa 5:20
136.4 C 224kPa 5:50
136.1 C 222kPa 6:20
136.3 C 223kPa 6:50
136.4 C 223kPa 7:20
136.4 C 223kPa 7:20
VENTING 7:20
AIR DRYING 8:12
CYCLE COMPLETE 8:41

Digital Signature #
CF3ADC4A2780348E
