
Midmark M11 - v1.0.0
Total Cycles: 1589

Name: RITTER 1
Sterilizer ID: V2178198
02/02/2022 01:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.1 F 0.0
2:00 156.5 F 2.9
4:00 171.0 F 6.7
6:00 205.7 F 11.1
8:00 215.2 F 4.7
10:00 229.7 F 9.8
12:00 235.1 F 8.5
14:00 241.7 F 12.0
16:00 239.5 F 10.1
18:00 252.9 F 16.6
20:00 264.1 F 23.3
22:00 271.2 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.5 F 28.9
2:00 271.5 F 28.2
3:00 272.2 F 28.9
4:00 272.2 F 28.7

Min 271.2 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.9 F 0.0
5:00 229.9 F 0.0
10:00 237.2 F 0.0
15:00 225.5 F 0.0
20:00 225.9 F 0.0
25:00 229.8 F 0.0
30:00 230.1 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 22:02
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:00:21
CYCLE COMPLETE

02/02/2022 02:11 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____