
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006125
15:39 15/08/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:10
PRESSURIZING 2:13
STERILIZING 3:13
136.0 C 220kPa 3:13
136.2 C 220kPa 3:43
136.3 C 223kPa 4:13
136.3 C 223kPa 4:43
136.2 C 224kPa 5:13
136.4 C 224kPa 5:43
136.3 C 221kPa 6:13
136.4 C 225kPa 6:43
136.4 C 225kPa 6:43
VENTING 6:44
AIR DRYING 7:39
CYCLE COMPLETE 7:43

Digital Signature #
563F9DC7BF4FCCF3
