
Midmark M11 - v1.0.0
Total Cycles: 3016

Name: RITTER 1
Sterilizer ID: V2178198
03/20/2023 07:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.1 F 0.0
2:00 79.9 F 0.0
4:00 86.8 F 0.0
6:00 115.6 F 0.8
8:00 145.4 F 2.4
10:00 177.6 F 4.9
12:00 194.1 F 8.4
14:00 209.5 F 12.6
16:00 219.3 F 5.7
18:00 229.9 F 9.8
20:00 235.8 F 10.1
22:00 239.4 F 11.0
24:00 235.7 F 8.6
26:00 249.8 F 15.1
28:00 261.2 F 21.5
30:00 271.6 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 271.9 F 28.5
2:00 272.4 F 29.0
3:00 271.7 F 28.7
4:00 271.8 F 28.4

Min 271.4 F 28.2
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.8 F 0.0
5:00 212.9 F 0.0
10:00 196.9 F 0.0
15:00 192.1 F 0.0
20:00 190.4 F 0.0
25:00 195.3 F 0.0
30:00 215.2 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 30:33
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:08:31
CYCLE COMPLETE

03/20/2023 08:42 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____