
Midmark M11 - v1.0.0
Total Cycles: 4775

Name: RITTER 1
Sterilizer ID: V2178198
07/29/2024 01:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 99.0 F 0.0
2:00 101.3 F 0.9
4:00 122.3 F 3.1
6:00 155.8 F 5.9
8:00 176.8 F 9.0
10:00 197.6 F 12.3
12:00 206.8 F 3.7
14:00 218.9 F 7.2
16:00 229.5 F 11.1
18:00 228.5 F 6.4
20:00 240.4 F 11.9
22:00 238.4 F 9.8
24:00 250.8 F 15.7
26:00 261.6 F 21.9
28:00 271.9 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.8 F 29.1
2:00 271.9 F 28.4
3:00 272.1 F 28.8
4:00 272.6 F 29.0

Min 271.5 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.0 F 0.0
5:00 214.2 F 0.0
10:00 202.9 F 0.0
15:00 198.6 F 0.0
20:00 221.1 F 0.0
25:00 222.3 F 0.0
30:00 228.5 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 28:36
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:07:06
CYCLE COMPLETE

07/29/2024 02:16 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____