

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001183  
15:04 21/07/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:39  
PRESSURIZING 2:46  
STERILIZING 3:56  
136.0 C 220kPa 3:56  
136.3 C 223kPa 4:26  
136.4 C 225kPa 4:56  
136.4 C 224kPa 5:26  
136.2 C 224kPa 5:56  
136.2 C 223kPa 6:26  
136.4 C 224kPa 6:56  
136.4 C 225kPa 7:26  
136.4 C 225kPa 7:26  
VENTING 7:26  
AIR DRYING 8:17  
CYCLE COMPLETE 19:36  
-----

Digital Signature #  
0360916D9B57ACB8  
-----