
Midmark M11 - v1.0.5
Total Cycles: 4214

Name: RITTER2
Sterilizer ID: V2313228
12/13/2024 04:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.1 F 0.0
2:00 164.1 F 2.3
4:00 178.0 F 5.9
6:00 199.2 F 10.3
8:00 221.5 F 6.4
10:00 229.8 F 8.7
12:00 238.6 F 13.6
14:00 239.7 F 10.8
16:00 236.4 F 9.0
18:00 250.1 F 15.3
20:00 260.8 F 21.5
22:00 270.7 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.7 F 29.8
2:00 272.7 F 29.3
3:00 271.5 F 28.9
4:00 272.4 F 29.7

Min 271.4 F 28.8
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.4 F 0.0
5:00 244.3 F 0.0
10:00 234.9 F 0.0
15:00 206.7 F 0.0
20:00 196.4 F 0.0
25:00 205.3 F 0.0
30:00 200.5 F 0.0
Drying Complete

FILLING: 1:43
HEATING: 23:49
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:03:19
CYCLE COMPLETE

12/13/2024 05:54 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____