

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4333

Name: RITTER 1  
Sterilizer ID: V2178198  
02/26/2024 12:59 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 123.9 F 0.0  
2:00 122.9 F 1.4  
4:00 142.2 F 3.9  
6:00 166.7 F 7.4  
8:00 203.3 F 11.8  
10:00 215.3 F 4.1  
12:00 227.1 F 8.1  
14:00 237.1 F 12.6  
16:00 235.2 F 8.7  
18:00 242.4 F 11.6  
20:00 243.9 F 12.1  
22:00 255.7 F 18.2  
24:00 266.6 F 24.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.3  
1:00 272.4 F 28.6  
2:00 272.0 F 28.8  
3:00 272.1 F 28.4  
4:00 272.2 F 29.0

Min 271.3 F 28.1  
Max 272.8 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.5 F 0.0  
5:00 242.4 F 0.0  
10:00 230.0 F 0.0  
15:00 223.5 F 0.0  
20:00 223.2 F 0.0  
25:00 224.0 F 0.0  
30:00 233.1 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 25:35  
STERILIZING: 4:00  
VENTING: 3:25  
DRYING: 30:00  
TOTAL CYCLE: 01:03:59  
CYCLE COMPLETE

02/26/2024 02:02 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_