
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.6uS / 0.3ppm
CYCLE NUMBER 004753
12:17 22/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:03
PRESSURIZING 2:05
STERILIZING 3:04
136.5 C 222kPa 3:04
136.3 C 225kPa 3:34
136.4 C 218kPa 4:04
136.4 C 224kPa 4:34
136.5 C 224kPa 5:04
136.4 C 224kPa 5:34
136.2 C 221kPa 6:04
136.4 C 220kPa 6:34
136.4 C 220kPa 6:34
VENTING 6:34
AIR DRYING 7:34
CYCLE COMPLETE 8:06

Digital Signature #
AE3E6637A38C48F3
