

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3944

Name: RITTER 1  
Sterilizer ID: V2178198  
10/25/2023 09:18 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.6 F 0.0  
2:00 71.2 F 0.0  
4:00 81.0 F 0.0  
6:00 113.3 F 0.0  
8:00 151.9 F 0.3  
10:00 178.9 F 3.5  
12:00 192.0 F 6.6  
14:00 209.5 F 9.9  
16:00 220.5 F 13.5  
18:00 223.9 F 7.0  
20:00 234.0 F 11.0  
22:00 231.6 F 7.2  
24:00 243.7 F 12.9  
26:00 241.7 F 11.1  
28:00 254.0 F 17.3  
30:00 264.7 F 23.7  
32:00 271.1 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.3  
1:00 272.5 F 29.0  
2:00 271.3 F 28.2  
3:00 272.4 F 29.1  
4:00 272.1 F 28.7  
  
Min 271.1 F 28.1  
Max 272.5 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 178.8 F 0.0  
5:00 199.8 F 0.0  
10:00 219.0 F 0.0  
15:00 215.2 F 0.0  
20:00 217.6 F 0.0  
25:00 223.0 F 0.0  
30:00 223.9 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 32:02  
STERILIZING: 4:00  
VENTING: 3:23  
DRYING: 30:00  
TOTAL CYCLE: 01:10:14  
CYCLE COMPLETE

10/25/2023 10:28 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_