
Midmark M11 - v1.0.0
Total Cycles: 3596

Name: RITTER 1
Sterilizer ID: V2178198
07/31/2023 04:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 136.4 F 0.0
2:00 138.0 F 0.0
4:00 146.4 F 0.5
6:00 180.0 F 3.9
8:00 204.2 F 7.8
10:00 213.7 F 12.6
12:00 224.6 F 6.8
14:00 236.6 F 11.6
16:00 234.1 F 8.5
18:00 242.1 F 11.0
20:00 245.5 F 13.0
22:00 257.4 F 19.3
24:00 268.1 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.4 F 28.8
2:00 271.2 F 28.2
3:00 272.6 F 29.0
4:00 271.6 F 28.3

Min 271.1 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.1 F 0.0
5:00 205.2 F 0.0
10:00 206.8 F 0.0
15:00 211.7 F 0.0
20:00 210.8 F 0.0
25:00 205.4 F 0.0
30:00 204.3 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 25:18
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:03:33
CYCLE COMPLETE

07/31/2023 05:10 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____