
Midmark M11 - v1.0.0
Total Cycles: 1653

Name: RITTER 1
Sterilizer ID: V2178198
02/22/2022 10:52 AM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 141.9 F 0.0
2:00 153.6 F 3.0
4:00 182.3 F 7.1
6:00 207.7 F 13.4
8:00 224.4 F 7.4
10:00 235.5 F 9.7
12:00 244.4 F 12.8
14:00 243.4 F 11.7
16:00 259.3 F 20.1
18:00 272.6 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.0
1:00 272.8 F 28.7
2:00 271.7 F 27.9
3:00 272.7 F 28.8

Min 271.7 F 27.8
Max 272.8 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 170.2 F 0.0
Dry Cycle Aborted

FILLING: 1:22
HEATING: 18:15
STERILIZING: 3:00
VENTING: 2:27
DRYING: 1:44
TOTAL CYCLE: 00:26:48

02/22/2022 11:18 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____