
Midmark M11 - v1.0.0
Total Cycles: 1676

Name: RITTER 1
Sterilizer ID: V2178198
02/28/2022 09:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 105.2 F 0.0
2:00 110.2 F 0.0
4:00 136.4 F 1.1
6:00 157.0 F 2.7
8:00 180.9 F 5.3
10:00 198.9 F 9.1
12:00 213.2 F 13.4
14:00 223.3 F 6.7
16:00 235.0 F 11.4
18:00 233.1 F 7.8
20:00 245.6 F 12.8
22:00 244.3 F 12.1
24:00 256.7 F 18.7
26:00 267.9 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.0
1:00 272.7 F 28.8
2:00 272.0 F 28.1
3:00 271.7 F 28.2
4:00 272.9 F 28.9

Min 271.5 F 27.9
Max 272.9 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 172.0 F 0.0
5:00 211.7 F 0.0
10:00 216.5 F 0.0
15:00 228.2 F 0.0
20:00 222.9 F 0.0
25:00 225.5 F 0.0
30:00 221.3 F 0.0

Drying Complete

FILLING: 0:56
HEATING: 27:21
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:05:23
CYCLE COMPLETE

02/28/2022 11:03 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____